

Evaluation of Laboratory Performance in MSCR Testing (T350/D7405) Using AASHTO re:source PSP Data

John J. Malusky, Program Manager,
AASHTO re:source Proficiency Sample Program

Initial Concerns:

- Laboratories are receiving satisfactory ratings (± 3 , ± 4 , ± 5 s) on percent recovery and J_{nr} values at 0.1 and 3.2 kPa, but receiving low ratings (0, ± 1 s, ± 2 s) on the percent differences (recovery and J_{nr}).

Sample 237				
Lab Data	Avg	1S	Z-Score	Rating
2.630	2.6246	0.2158	0.03	5

Creep and Recovery (MSCR)
Creep Compliance at 3.2 kPa, $J_{nr3.2}$ (0.001 significant figure)
ogram | View Performance Chart

Sample 237				
Lab Data	Avg	1S	Z-Score	Rating
3.170	3.0772	0.2364	0.39	5

Creep and Recovery (MSCR)
Percent Difference of Non-recoverable Creep Compliance, $J_{nr-diff}$ (0.01 significant figure)
ogram | View Performance Chart

Sample 237				
Lab Data	Avg	1S	Z-Score	Rating
20.40	16.577	1.476	2.59	1

Sample 238				
Lab Data	Avg	1S	Z-Score	Rating
2.440	2.6047	0.2109	-0.78	-5

Percent Difference of Non-recoverable Creep Compliance, $J_{nr-diff}$ (0.01 significant figure) - TP70/D7405

Sample 238				
Lab Data	Avg	1S	Z-Score	Rating
3.000	3.0504	0.2364	-0.21	-5

Percent Difference of Non-recoverable Creep Compliance, $J_{nr-diff}$ (0.01 significant figure) - TP70/D7405

Sample 238				
Lab Data	Avg	1S	Z-Score	Rating
23.05	16.556	1.427	4.55	0

Evaluation of the Issue:

- ▶ From the initial feedback and comments we determined that this was an isolated event happening in one PSP round. Caused by the difference in values between the “+5s and the -5s”.
- ▶ Looking back on our first thought - “difference between a +5 and a -5”.
 - ▶ It doesn't matter where the data falls when calculating a % difference.

$$J_{\%diff} = \frac{[J_{nr3.2} - J_{nr0.1}] \times 100}{J_{nr0.1}}$$

Lab 1		
0.1 kPa	3.2 kPa	% Diff
40	60	50

-5 +5 +5

Lab 2		
0.1 kPa	3.2 kPa	% Diff
30	45	50

-4 -5 +5

Lab 3		
0.1 kPa	3.2 kPa	% Diff
10	15	50

-2 -2 +5

Lab 4		
0.1 kPa	3.2 kPa	% Diff
60	90	50

+5 +2 +5

Statistically Significant Data PGB Rounds 241/242 (64-28p)

- ▶ Out of the six reporting parameters in T350/D7405, statistical differences existed between manufacturers (A, B, & C) for these four test parameters:
 - ▶ % Recovery at 0.1 kPa (A - B)
 - ▶ % Difference in Recovery (A - B)
 - ▶ J_{nr} at 0.1 kPa (A - B)
 - ▶ % Difference in J_{nr} (A - B - C)

Outcome:

- ▶ We will continue to solicit for test data for all reporting parameters in the MSCR (T350/D7405).
- ▶ Administrative Task Group has been informed of the situation.
 - ▶ AAP's proposed to the ATG is to **not** evaluate % difference in recovery and % difference in J_{nr} for accreditation purposes.
 - ▶ Still evaluate data for % recovery and J_{nr} values at 0.1 and 3.2 kPa, respectively.
- ▶ Continue to evaluate the data after each PSP round and look for issues (check model and software version).

Results from PGB 243/244


(PG 64-22)

(evaluation using Welch-t)


- ▶ Statistical significance exists between manufacturers for the following parameters:
 - ▶ % recovery at 0.1 kPa (all manufacturers)
 - ▶ % recovery at 3.2 kPa (all manufacturers)
 - ▶ % difference in recovery (all manufacturers)
 - ▶ % difference in jnr (all manufacturers)

Scatter Plots

(Percent Recovery at 0.1 kPa)




Percent Recovery at 3.2kPa



Bi-modal distribution


Ratings were suppressed

Percent Difference in Recovery



Bi-modal distribution
Ratings suppressed

Percent Difference in jnr



Discussion:

- ▶ Contacted DSR Manufacturers to cross reference the reported versions.
 - ▶ Communication indicates that laboratories are not certain on what type of software they have.
 - ▶ DSR manufacturers are reaching out to customers to ensure that software is being updated to the most current versions.
- ▶ AASHTO re:source Assessments:
 - ▶ Identifying devices w/o most current software.
 - ▶ Assessors are looking for the data to determine if conditioning cycles are being used.
 - ▶ Implemented in 2014 - tour cycle is close to 30 months (6 month lag)

Options

- ▶ Collect data based off of the correct software versions.
 - ▶ Be more clear in specialized sample round instructions.
- ▶ New RTFO sample vs. tested RTFO DSR sample with “rest” period (AASHTO and ASTM allow both)
- ▶ Revise the standards to require most current version of software from the manufacturer
- ▶ Any suggestions?

Developing Precision Estimates


- ▶ Manner of expression of estimates (AASHTO and ASTM):
 - ▶ Standard deviation with 95% confidence interval
 - ▶ Coefficient of Variation expressed as a percentage
- ▶ Regression analysis:
 - ▶ Plot sample averages vs. the standard deviation and analyze with regression
 - ▶ Evaluate the points
 - ▶ Evaluate the r^2 value
- ▶ Determine the manner of expression
 - ▶ High r^2 = use % CV, low r^2 = use 1s

Binder Rounds and Type

Sample ID	No. of Participants	PG Grade	MSCR Grade
229 & 230	150	70-28 (p)	70-28(H)
233 & 234	163	82-22 (p)	FAILED*
235 & 236	181	58-28	58-28(S)
237 & 238	181	70-22	70-22(S)
239 & 240	196	64-22	64-22(H)
241 & 242	207	58-28 (p)	58-28(H)
243 & 244	209	64-22	64-22(S)
* % Diff in Jnr was >75%			




Estimates From MSCR Test Parameters

► Percent Recovery at 0.1 kPa



	1s	d2s
Single Operator	0.72	2.0
Inter-laboratory	3.14	8.8


► Percent Recovery at 3.2 kPa



	1s	d2s
Single Operator	0.85	2.4
Inter-laboratory	2.68	7.5


	1s%	
Single Operator	5.99	16.8
Inter-laboratory	23.90	66.9

► Percent Difference in Recovery




	1s	d2s
Single Operator	1.86	5.2
Inter-laboratory	7.90	22.1

► Jnr at 0.1 kPa



	1s%	d2s%
Single Operator	3.61	10.1
Inter-laboratory	8.99	25.2


► Jnr at 3.2 kPa



	1s	d2s
Single Operator	0.08	0.2
Inter-laboratory	0.23	0.6



	1s%	d2s%
Single Operator	3.69	10.3
Inter-laboratory	8.56	24.0

► Percent Difference in Jnr



	1s%	d2s%
Single Operator	4.6	12.9
Inter-laboratory	15.6	43.7

	1s	d2s
Single Operator	3.36	9.4
Inter-laboratory	24.67	69.1



Results and Conclusion

- ▶ Evaluation of data sets is not clear:
 - ▶ Use 1s or %CV appears to be dependent on test parameter and on the material (modified vs. unmodified)
- ▶ Combination of 1s or %CV may be needed for different materials

Thank You!

John J. Malusky

Program Manager, Proficiency Sample Program



Email: jmalusky@ashtoresource.org

Direct: 240-436-4825

Main: 240-436-4900

Website: www.ashtoresource.org

AASHTO re:source (formerly AMRL)

4441 Buckeystown Pike

Suite A

Frederick, MD 21704